

ISET s.r.l.

Sede Legale e Uffici Via Donatori di sangue, 9 - 46024 Moglia (MN) Tel. e fax +39 (0)376 598963 iset@iset-italia.com www.iset-italia.eu Cap. soc. i.v. Cod. Fisc. e P.IVA Reg. Imprese

€ 10.200,00 02 332 750 369 02 332 750 369 MN 0221098

CERTIFICATE OF COMPLIANCE

Certificado de Conformidade - Сертификат соответствия - Konformitätserklärung

1) APPLICANT:

Flo-Instruments Co., Ltd 1507, Building F, Bantian International Center, No.5 Huancheng South Road, Bantian Street, Longgang District, Shenzhen, 518129 **2) CERTIFICATE NO.:** IT020945UT170309

TCF(S) NO.: SZUTTCF-LVD/EMC

3) WITH REFERENCE TO EC DIRECTIVE APPLIED:

Low Voltage Directive 2014/35/EU Electromagnetic Compatibility 2014/30/EU

4) CERTIFICATION ISET MARK:



HARMONIZED STANDARDS APPLIED:

EN 61010-1:2010,EN 61326-1:2013, EN 61000-3-2:2014,EN 61000-3-3:2013

5) PRODUCT CHARACTERISTICS: Ultrasonic Flowmeter

MODEL(S): ST501, ST502, TM601, TM602, MU801 Plus, PH301, TM601S, TM605, SE603, SE509

REMARK: The verification has been carried out on voluntary application of the manufacturer based only on the documents prepared and provided by the manufacturer itself. The product(s) satisfies the requirements of the Certification Mark of ISET according to the ISET regulations. The manufacturer is responsible to maintain the internal production control to ensure the compliance of the product. ISET declines any liability with reference to any other noncompliance of documents, product or test report that have been submitted to evaluation. However marking and EC declaration are duties of the manufacturer before putting into service of its product(s) on market as well as the verification doesn't imply assessment of the production of the product(s). The manufacturer is liable to take all the necessary actions required by the applicable directives&procedures before putting CE mark on the product(s). The manufacturer should be responsible for the results caused by improperly use this verification.

6) DATE OF ISSUE: 09/03/2020 DATE OF EXPIRE: 08/03/2025

